



# Serinus® Cal 3000

Gas Dilution Calibrator with Internal Ozone Generator & Ozone Photometer (ozone transfer standard)

Reliability, accuracy and precision are hallmarks of Acoem's product range. It's why Acoem's Calibrators have long been trusted worldwide.

The Serinus Cal 3000\* includes an ozone photometer for the accurate creation and delivery of ozone concentrations when calibrating ozone analysers. The photometer accurately and continuously measures the ozone concentration to control output of the internal ozone generator, providing an accurate ozone source for routine ozone calibration.

<sup>\*</sup> Acoem Serinus Cal 3000 formerly known as Ecotech Serinus Cal 3000

#### **Features**

- Serinus Cal user interface makes all functions simple & intuitive
- In-built photometer & ozone generator for use as a Level 3 ozone transfer standard
- Ozone photometer based on the reliable & proven technology used in the Serinus 10 gas analyser.

### Dilution & span flows

Dilution gas inputs: 1 standard 100 - 200 kPa (g) (2 optional)

Source gas inputs: 4 standard 100 - 200 kPa (g) (8 optional)

**Dilution mass flow** 

controller: 10 SLPM, 0 Deg, 1 ATM (std),

1 SLPM, 2 SLPM, 5 SLPM or 20 SLPM (optional),

2nd MFC\* (optional)

Source mass flow

controller: 50 SCCM, 0 Deg, 1 ATM (std),

10 SCCM, 20 SCCM, 100 SCCM, 500 SCCM or 1 SLPM,

2 SLPM (optional), 2nd MFC\* (optional)

Flow accuracy

(constant temp):Within 1 % of full scaleFlow repeatability:Within 0.15 % of full scaleLinearity:Within 0.15 % of full scale

Operating gas pressure: 100 - 200 kPa

Zero drift: < 0.6 % per year

Response time: < 5 seconds

Output manifold: 4 output ports standard

Dilution ratio: Variable 10:1 to 2000:1 (std configuration)

**Case dimensions** 

rack length: 597 mm (23.5") (front to rear)

**Total length** 

(with latch release):638 mm (25.1")Chassis width:418 mm (16.5")Front panel width:429 mm (16.9")

Chassis height: 163 mm / uses 4RU (6.4")

Front panel height: 175 mm (6.9")

Weight: 23.8 kg



 $<sup>^{*}</sup>$  Additional source MFC reduces available source ports by 1 and results in no analog output being available

#### Communication

**User interface:** 

Via front panel keypad or computer

**Programmable** calibrations:

16 separate programmable sequences

32 separate programmable points

**Analog output:** Voltage output of 0 to 5 V, with menu selectable

zero off set of 0, 5 or 10 %

**Analog input:** Three analog voltage inputs (0 - 5 VDC) CAT I rated

**Digital output:** RS232 port #1: Normal digital communication

RS232 port #2: Multidrop port used for multiple

analyser connections on a single RS232 USB port connection on rear panel

25 pin connector with discrete status &

user control

USB stick memory (front panel) for data logging,

event logging & parameter storage

8 Digital Outputs, open collector max 400 mA

each at 12 VDC (max total output 2A)

8 Digital Inputs, 0 - 5 VDC

CAT I rated

1 Diluent Control, + 12 V output.

#### Power

Operating voltage: 100 - 240 V VAC 50 / 60 Hz (autoranging) Power consumption: 165 VA maximum 95 VA after warm-up

Operating conditions

ambient temperature

0 - 45 °C (32 - 104 °F), 20 - 35 °C range:

for optimum performance

Pressure: Maximum altitude: 3000 m above sea level.



## Ozone generator

Output concentration: 3 ppb to 5000 ppb

Flow rate: Variable dependent on Dilution Mass Flow Controller installed

Repeatability: < 1 % short term (24 hours) 5 % long term at constant

temperature & humidity.

#### **Photometer**

Range: 0-20 ppm

Precision: 0.5 ppb or 0.2 % of reading, whichever is greater

Linearity: < 1 % of full scale

Noise at zero: < 0.25 ppb

Response time: 30 seconds to 95 %

Zero drift: Temperature: 1.0 ppb per °C

24 hours: < 0.3 ppb

7 days: < 0.3 ppb

Span drift: Temperature: 0.1 % per °C

7 days: 0.5 % of reading.





Specifications subject to change without notice. Images used are for illustrative purposes only. All trademarks and registered trademarks are the property of their respective owners.

© 2022 Acoem and all related entities. All rights reserved. 20221019